

AMPri Handelsgesellschaft mbH Benzstraße 16 21423 Winsen (Luhe) - Germany Telefon: +49 (0)4171 / 8480-0 Homepage: www.ampri.de e-mail: info@ampri.de

| Technical | Data She | et | | | | | |
|--|--|---|---------------------------------------|--|---|--|--|
| Article-No.: | 01197 | | | | | | |
| Description: | | SIC-PLUS nination glove erile, powder f | froo | | | | |
| | blue, non st | erne, powder i | liee | | Constant of the second s | | |
| THICKNESS* cuff | palm fir | ngertips | | | | | |
| 0,04 mm | | ,05 mm | | | | | |
| PRODUCT DESCRIP | TION | | | | | | |
| material | Latex | ☑ Nitrile | 🗆 Vinyl | Vinyl-Nitrile- mixture | Polyethy-lene(PE) | 🗆 ТРЕ | Cotton |
| colour characteristics | whiteprepowdered | ✓ blue✓ powderfree | black sterile | ☐ mint☑ non sterile | purple ambidex- trave | mix fits hand- | bordeaux Aloe Vera inner conting |
| surface | ✓ textured | not textured | embossed | | trous | specific | inner coating |
| SIZES | | | | | | | |
| | XS (5-6) | S (6-7) | M (7-8) | L (8-9) 110 ± 10 mm | XL (9-10) 115 ± 10 mm | XXL (10-11) | XXXL (11-12) |
| width length | ≤ 80 mm ≥ 240 mm | 80 ± 10 mm ≥ 240 mm | 95 ± 10 mm ≥ 240 mm | 110 ± 10 mm ≥ 240 mm | 115 ± 10 mm ≥ 240 mm | - | - |
| REGULATORY AFFA | | - | u u u u u u u u u u u u u u u u u u u | | - | | |
| PPE-Regulation (EU) 2016/425 | Category I | Category II | Category III | no PPE-article | | | ~ |
| MD-Regulation (EU) 2017/745 | Class I | Class II a | Class III | □ sterile | measuring function | no medical device | CE |
| Food Contact (EG) 1935/2004 | ☑ acidic foods | aqueous foods | ✓ fatty foods | alcoholic foods | ✓ dry foods | not approved for food- contact | זיא |
| STANDARDISATION | N | | | | | | |
| EN 388 Mechanical Risks | abrasion resistance | blade cut resistance Coupe-Test | tear resistance | puncture resistance | blade cut resistance TDM-Test | impact test | |
| Level | not applicable | couperrest | | | 1 Divi-1est | | |
| EN 374-1 | | che | mical | | code letter | | |
| Chemical Risks | Sodium hydroxide | | | | К | | ISO 374-1/Type B |
| EN 374-4 Degradation | Hydrogen Peroxide 30% Formaldehyde 37% | | | | | P T | |
| EN 374-5 microorganism | The glove is tight against microorganisms (viral, bacteria and fungi). Test according to ISO 16604 - method B. | | | | | | KPT |
| tightness EN ISO 21420 | The glove meets the requirements according to EN ISO 21420 | | | | | ~ 11105 | |
| protective gloves | | | | | | | |
| EN 455 medical gloves for single use | The glove meets the requirements according to EN 455-1, EN 455-2, EN 455-3, EN 455-4. Force at break \ge 3.6 N. | | | | | | EN 455 |
| EN 455-1 freedom from holes | The glove has an AQL < 1.5 in regards to the water retention test (sampling inspection in acc. to ISO 2859-1, general Inspection Level I) | | | | | | AQL (1.5) |
| EN 16350 electrostatic properties | not applicable | | | | | | |



AMPri Handelsgesellschaft mbH Benzstraße 16 21423 Winsen (Luhe) - Germany Telefon: +49 (0)4171 / 8480-0 Homepage: www.ampri.de e-mail: info@ampri.de

Technical Data Sheet

Article-No.: 01197

Description:

BLUE BASIC-PLUS Nitrile examination glove blue, non sterile, powder free

| LOGISTIC DATA SUBPACKING | |
|---------------------------|---------------|
| generell information | |
| material | carton |
| pieces per subpacking | 200 |
| GTIN subpacking size XS | 4044941005300 |
| GTIN subpacking size S | 4044941005317 |
| GTIN subpacking size M | 4044941005324 |
| GTIN subpacking size L | 4044941005331 |
| GTIN subpacking size XL | 4044941005348 |
| GTIN subpacking size XXL | - |
| GTIN subpacking size XXXL | - |
| PZN subpacking size XS | 15577716 |
| PZN subpacking size S | 15577685 |
| PZN subpacking size M | 15577662 |
| PZN subpacking size L | 15577679 |
| PZN subpacking size XL | 15577691 |
| PZN subpacking size XXL | - |
| PZN subpacking size XXXL | - |
| measures & size | |
| length | 240 mm |
| width | 128 mm |
| heigth | 80 mm |
| weights | |
| size | gross weight* |
| XS | 580 g |
| S | 640 g |
| Μ | 700 g |
| L | 760 g |
| XL | 820 g |
| XXL | - |
| XXXL | - |

| LOGISTIC DATA PALETTE | | |
|-----------------------|---------------|--|
| general information | | |
| kind of palett | euro-palette | |
| measures & size | | |
| cartons per layer | 6 | |
| layers per palette | 7 | |
| heigth of the palette | 190 cm | |
| weights | | |
| size | gross weight* | |
| XS | 290 kg | |
| S | 315 kg | |
| Μ | 340 kg | |
| L | 365 kg | |
| XL | 390 kg | |
| XXL | - | |
| XXXL | - | |



| LOGISTIC DATA OUTER PACKING | |
|-------------------------------|---------------|
| generell information | |
| material | carton |
| subpackings per outer packing | 10 |
| GTIN outer packing size XS | 4044941005355 |
| GTIN outer packing size S | 4044941005362 |
| GTIN outer packing size M | 4044941005379 |
| GTIN outer packing size L | 4044941005386 |
| GTIN outer packing size XL | 4044941005393 |
| GTIN outer packing size XXL | - |
| GTIN outer packing size XXXL | - |
| PZN outer packing size XS | - |
| PZN outer packing size S | - |
| PZN outer packing size M | - |
| PZN outer packing size L | - |
| PZN outer packing size XL | - |
| PZN outer packing size XXL | - |
| PZN outer packing size XXXL | - |
| measures & size | |
| length | 415 mm |
| width | 265 mm |
| heigth | 250 mm |
| weights | |
| size | gross weight* |
| XS | 6.300 g |
| S | 6.900 g |
| Μ | 7.500 g |
| L | 8.100 g |
| XL | 8.700 g |
| XXL | - |
| XXXL | - |



AMPri Handelsgesellschaft mbH Benzstraße 16 21423 Winsen (Luhe) - Germany Telefon: +49 (0)4171 / 8480-0 Homepage: www.ampri.de e-mail: info@ampri.de

Technical Data Sheet

Article-No.: 01197

Description:

BLUE BASIC-PLUS Nitrile examination glove blue, non sterile, powder free



WARNINGS AND SAFETY INFORMATION storage / expiry Store gloves in original packaging in a cool and dry place without additional weight, protect from direct sunlight. Do not store near ozone date sources (laser printers, copiers). The actual expiry time in use cannot be specified in general terms, as it depends on the general conditions of use. An individual risk assessment must be carried out in each case. The expiry date - valid for proper storage - is stated on the packaging. Always use protective gloves only for the intended use and in the correct size. A check/risk assessment must be carried out to ensure that use and control the gloves are suitable for the intended use, as the conditions at the workplace may deviate from those of the type test depending on temperature, abrasion and degradation. Breakthrough times and permeation levels are based on laboratory measurements and are determined using samples taken from the palm of the hand. The actual duration of protection of a glove with a specific substance can vary significantly due to the conditions of use (temperature, abrasion, stretching). In the case of aggressive chemicals, degradation (change in mechanical properties) can be an important factor to consider when selecting chemical-resistant gloves. This information does not reflect the actual duration of protection in the workplace and the distinction between mixtures and pure chemicals. The chemical resistance was determined under laboratory conditions only on the basis of samples from the palm and refers only to the chemicals tested. The situation may be different if the chemical is used in a mixture. The penetration resistance was evaluated under laboratory conditions and refers only to the tested specimen. The degradation results according to EN ISO 374-4 show the change in puncture resistance of the gloves after exposure to the tested chemical. Before use, the gloves must be checked for holes or damage. disposal Used gloves must be disposed of after contact with chemicals in accordance with the disposal regulations for the chemical and the regulations of the local waste disposal company. Unused gloves can be disposed of with household waste. disinfection Disinfection is not intended for these gloves and is the responsibility of the user. warnings/ Protective gloves are intended for single use only.

| allergy | This product contains dithiocarbamates, which may cause allergic reactions |
|--|--|
| donning and doffing instructions | |

*slight deviations possible due to standard tolerances

changes and errors excepted

rev-no.: 2025-01 date 16.06.2025