

## Technical Data Sheet

Article-No.: **01194**

Description: **MED-COMFORT Blue Ultra 300**  
Nitrile examination glove  
blue, non sterile, powder free





THICKNESS*		
cuff	palm	fingertips
0,09 mm	0,13 mm	0,15 mm

PRODUCT DESCRIPTION							
material	<input type="checkbox"/> Latex	<input checked="" type="checkbox"/> Nitrile	<input type="checkbox"/> Vinyl	<input type="checkbox"/> Vinyl-Nitrile-mixture	<input type="checkbox"/> Polyethy-lene (PE)	<input type="checkbox"/> TPE	<input type="checkbox"/> cotton
colour	<input type="checkbox"/> white	<input checked="" type="checkbox"/> blue	<input type="checkbox"/> black	<input type="checkbox"/> mint	<input type="checkbox"/> purple	<input type="checkbox"/> mix	<input type="checkbox"/> bordeaux
characteristics	<input type="checkbox"/> prepowdered	<input checked="" type="checkbox"/> powderfree	<input type="checkbox"/> sterile	<input checked="" type="checkbox"/> non sterile	<input checked="" type="checkbox"/> ambidextrous	<input type="checkbox"/> fits hand-specific	<input type="checkbox"/> Aloe Vera inner coating
surface	<input checked="" type="checkbox"/> textured	<input type="checkbox"/> not textured	<input type="checkbox"/> embossed				

SIZES							
	XS (5-6)	S (6-7)	M (7-8)	L (8-9)	XL (9-10)	XXL (10-11)	XXXL (11-12)
width	-	80 ± 10 mm	95 ± 10 mm	110 ± 10 mm	115 ± 10 mm	125 ± 10 mm	-
length	-	≥ 300 mm	≥ 300 mm	≥ 300 mm	≥ 300 mm	≥ 300 mm	-

REGULATORY AFFAIRS							
PPE-Regulation (EU) 2016/425	<input type="checkbox"/> Category I	<input type="checkbox"/> Category II	<input checked="" type="checkbox"/> Category III	<input type="checkbox"/> no PPE-article			CE
MD-Regulation (EU) 2017/745	<input checked="" type="checkbox"/> Class I	<input type="checkbox"/> Class II a	<input type="checkbox"/> Class III	<input type="checkbox"/> sterile	<input type="checkbox"/> measuring function	<input type="checkbox"/> no medical device	
Food Contact (EG) 1935/2004	<input checked="" type="checkbox"/> acidic foods	<input checked="" type="checkbox"/> aqueous foods	<input checked="" type="checkbox"/> fatty foods	<input checked="" type="checkbox"/> alcoholic foods	<input checked="" type="checkbox"/> dry foods	<input type="checkbox"/> not approved for food-contact	RF

STANDARDISATION							
EN 388 Mechanical Risks	abrasion resistance	blade cut resistance Coupe-Test	tear resistance	puncture resistance	blade cut resistance TDM-Test	impact test	
Level	not applicable						

EN 374-1	chemical	code letter	 ISO 374-1/Type B
Chemical Risks	Sodium hydroxide 40%	K	
	Hydrogen Peroxide 30%	P	
EN 374-4	Formaldehyde 37%	T	
Degradation			
EN 374-5 microorganism tightness	The glove is tight against microorganisms (viral, bacteria and fungi). Test according to ISO 16604 - method B.		 EN ISO 374-5:2016 VIRUS

EN ISO 21420 protective gloves	The glove meets the requirements according to EN ISO 21420						
--------------------------------	--	--	--	--	--	--	--

EN 455 medical gloves for single use	The glove meets the requirements according to EN 455-1, EN 455-2, EN 455-3, EN 455-4. Force at break ≥ 3.6 N.						EN 455
--------------------------------------	---	--	--	--	--	--	--------

EN 455-1 freedom from holes	The glove has an AQL < 1.5 in regards to the water retention test ( sampling inspection in acc. to ISO 2859-1, general Inspection Level I)						AQL 1.5
-----------------------------	--	--	--	--	--	--	---------

EN 16350 electrostatic properties	not applicable						
-----------------------------------	----------------	--	--	--	--	--	--

## Technical Data Sheet

Article-No.: **01194**

Description: **MED-COMFORT Blue Ultra 300**  
Nitrile examination glove  
blue, non sterile, powder free



LOGISTIC DATA SUBPACKING	
generell information	
material	carton
pieces per subpacking	100
GTIN subpacking size XS	-
GTIN subpacking size S	4044941009964
GTIN subpacking size M	4044941009971
GTIN subpacking size L	4044941009988
GTIN subpacking size XL	4044941040519
GTIN subpacking size XXL	4044941010526
GTIN subpacking size XXXL	-
PZN subpacking size XS	-
PZN subpacking size S	15570341
PZN subpacking size M	15570335
PZN subpacking size L	15570329
PZN subpacking size XL	15570385
PZN subpacking size XXL	15570364
PZN subpacking size XXXL	-
measures & size	
length	280 mm
width	135 mm
height	70 mm
weights	
size	gross weight*
XS	-
S	710 g
M	760 g
L	800 g
XL	840 g
XXL	880 g
XXXL	-

LOGISTIC DATA PALETTE	
general information	
kind of palett	euro-palett
measures & size	
cartons per layer	8
layers per palette	6
height of the palette	192 cm
weights	
size	gross weight*
XS	-
S	390 kg
M	414 kg
L	433 kg
XL	452 kg
XXL	471 kg
XXXL	-

LOGISTIC DATA OUTER PACKING	
generell information	
material	carton
subpackings per outer packing	10
GTIN outer packing size XS	-
GTIN outer packing size S	4044941010533
GTIN outer packing size M	4044941010540
GTIN outer packing size L	4044941010557
GTIN outer packing size XL	4044941010564
GTIN outer packing size XXL	4044941010571
GTIN outer packing size XXXL	-
PZN outer packing size XS	-
PZN outer packing size S	-
PZN outer packing size M	-
PZN outer packing size L	-
PZN outer packing size XL	-
PZN outer packing size XXL	-
PZN outer packing size XXXL	-
measures & size	
length	390 mm
width	285 mm
height	295 mm
weights	
size	gross weight*
XS	-
S	7.600 g
M	8.100 g
L	8.500 g
XL	8.900 g
XXL	9.300 g
XXXL	-

## Technical Data Sheet

Article-No.: **01194**  
Description: **MED-COMFORT Blue Ultra 300**  
**Nitrile examination glove**  
**blue, non sterile, powder free**



### WARNINGS AND SAFETY INFORMATION

<b>storage / expiry date</b>	Store gloves in original packaging in a cool and dry place without additional weight, protect from direct sunlight. Do not store near ozone sources (laser printers, copiers). The actual expiry time in use cannot be specified in general terms, as it depends on the general conditions of use. An individual risk assessment must be carried out in each case. The expiry date - valid for proper storage - is stated on the packaging.
<b>use and control</b>	Always use protective gloves only for the intended use and in the correct size. A check/risk assessment must be carried out to ensure that the gloves are suitable for the intended use, as the conditions at the workplace may deviate from those of the type test depending on temperature, abrasion and degradation. Breakthrough times and permeation levels are based on laboratory measurements and are determined using samples taken from the palm of the hand. The actual duration of protection of a glove with a specific substance can vary significantly due to the conditions of use (temperature, abrasion, stretching). In the case of aggressive chemicals, degradation (change in mechanical properties) can be an important factor to consider when selecting chemical-resistant gloves. This information does not reflect the actual duration of protection in the workplace and the distinction between mixtures and pure chemicals. The chemical resistance was determined under laboratory conditions only on the basis of samples from the palm and refers only to the chemicals tested. The situation may be different if the chemical is used in a mixture. The penetration resistance was evaluated under laboratory conditions and refers only to the tested specimen. The degradation results according to EN ISO 374-4 show the change in puncture resistance of the gloves after exposure to the tested chemical. Before use, the gloves must be checked for holes or damage.
<b>disposal</b>	Used gloves must be disposed of in accordance with the disposal regulations of the local waste disposal company. Unused gloves can be disposed of with household waste.
<b>disinfection</b>	Disinfection is not intended for these gloves and is the responsibility of the user.
<b>warnings/ allergy information</b>	Protective gloves are intended for single use only. This product contains dithiocarbamates, which may cause allergic reactions
<b>donning and doffing instructions</b>	

\*slight deviations possible due to standard tolerances

rev-no.: 2025-01  
date 16.06.2025

changes and errors excepted