



Benzstraße 16 21423 Winsen (Luhe) - Germany

Telefon: +49 (0)4171 / 8480-0 Homepage: www.ampri.de e-mail: info@ampri.de

## **Technical Data Sheet**


Article-No.: 01185

EN

Description: STYLE STRAWBERRY

Nitrile examination glove


light pink, non sterile, powder free






colour	colour	e	mi	xture (PE) int	☐ mix ☐ fits hand- specific	☑ light pink ☐ biodegra- dable
white	characteristics	owdered	□ black □ mi □ sterile □ no □ embossed  M (7-8) 1 95 ± 10 mm 110	nt	☐ fits hand- specific	biodegra- dable
characteristics	characteristics	owdered	□ sterile □ no □ embossed  M (7-8) 95 ± 10 mm 110	n sterile  ambidex- trous	☐ fits hand- specific	biodegra- dable
SIZES    XS (5-6)   S (6-7)   M (7-8)   L (8-9)   XL (9-10)   XXL (10-11)   XXXL (11-12)	surface    SIZES  width   ≤ 80 length   ≥ 240  REGULATORY AFFAIRS  PPE-Regulation    Cate	ured □ not textured  (5-6) S (6-7) 0 mm 80 ± 10 mm	☐ embossed ☐ M (7-8) ☐ B 110 ☐ M (7-8) ☐ B 110 ☐ M 11	trous  L (8-9) XL (9-10)	specific	dable
SIZES    XS (5-6)   S (6-7)   M (7-8)   L (8-9)   XL (9-10)   XXL (10-11)   XXXL (11-12)	SIZES	(5-6) S (6-7) 0 mm 80 ± 10 mm	<b>M (7-8)</b> 95 ± 10 mm 110	L (8-9) XL (9-10)		
SIZES    XS (5-6)   S (6-7)   M (7-8)   L (8-9)   XL (9-10)   XXL (10-11)   XXXL (11-12)	SIZES    XS (   width   ≤ 80     length   ≥ 240     REGULATORY AFFAIRS   PPE-Regulation   □ Cate	(5-6) S (6-7) 0 mm 80 ± 10 mm	<b>M (7-8)</b> 95 ± 10 mm 110	, ,	XXL (10-11)	XXXL (11-12)
XS (5-6)   S (6-7)   M (7-8)   L (8-9)   XL (9-10)   XXL (10-11)   XXXL (11-12)	xs (           width         ≤ 80           length         ≥ 240           REGULATORY AFFAIRS           PPE-Regulation         □ Cate	0 mm 80 ± 10 mm	95 ± 10 mm 110	, ,	XXL (10-11)	XXXL (11-12)
XS (5-6)   S (6-7)   M (7-8)   L (8-9)   XL (9-10)   XXL (10-11)   XXXL (11-12)	xs (           width         ≤ 80           length         ≥ 240           REGULATORY AFFAIRS           PPE-Regulation         □ Cate	0 mm 80 ± 10 mm	95 ± 10 mm 110	, ,	XXL (10-11)	XXXL (11-12)
XS (5-6)   S (6-7)   M (7-8)   L (8-9)   XL (9-10)   XXL (10-11)   XXXL (11-12)	xs (           width         ≤ 80           length         ≥ 240           REGULATORY AFFAIRS           PPE-Regulation         □ Cate	0 mm 80 ± 10 mm	95 ± 10 mm 110	, ,	XXL (10-11)	XXXL (11-12)
Second Contact   Gegin Second Contact   Ge	width ≤ 80 length ≥ 240  REGULATORY AFFAIRS  PPE-Regulation □ Cate	0 mm 80 ± 10 mm	95 ± 10 mm 110	, ,	VVF (10-11)	VVVF (11-15)
Pength   240 mm   2	length ≥ 240   REGULATORY AFFAIRS   PPE-Regulation					_
REGULATORY AFFAIRS  PPE-Regulation (EU) 2016/425  MD-Regulation (EU) 2017/745  Food Contact (EG) 1935/2004  STANDARDISATION  EN 388 Mechanical Risks  Risks  Risks  Resultation Category II Category III On PPE-article (EG) 1935/2004  Category III On PPE-article (EG) 1935/2004  Class II Oclass III O	REGULATORY AFFAIRS  PPE-Regulation		1 440 HHH 1 47		-	-
PPE-Regulation (EU) 2016/425  MD-Regulation (EU) 2017/745  Food Contact (EG) 1935/2004  Class II a Class III   sterile   measuring function   device    Food Contact (EG) 1935/2004  Class II a class III   sterile   measuring function   device    Food Contact (EG) 1935/2004  Foods  F	PPE-Regulation			,		
Class   Clas	•					
Code	(FII) 204 C (42F	gory I Category II	☑ Category III ☐ no	PPE-article		
Code	_	-	Class III at			CE
Food Contact (EG) 1935/2004    acidic foods   aqueous foods   fatty foods   alcoholic foods   dry foods   not approved for food-contact     STANDARDISATION   EN 388 Mechanical Risks   resistance   res		S I ☐ Class II a	□ Class III □ ste	J		
Foods   Foods   For food-contact   For food-contact		is foods 🗸 agusous	✓ fatty foods ✓ alv		i	
STANDARDISATION  EN 388 Mechanical abrasion blade cut tear resistance puncture blade cut impact test resistance resistance resistance resistance TDM-Test  Level not applicable chemical code letter			,			U"
STANDARDISATION  EN 388 Mechanical abrasion blade cut tear resistance puncture blade cut impact test resistance resistance resistance resistance TDM-Test  Level not applicable chemical code letter	(EG) 1953/2004	10005	100	Jus		JCI
EN 388 Mechanical abrasion blade cut tear resistance puncture blade cut impact test  Risks resistance resistan					contact	
Risks resistance resis	STANDARDISATION					
Coupe-Test TDM-Test Level not applicable Chemical code letter	EN 388 Mechanical abra	asion blade cut	tear resistance pu	ıncture blade cut	impact test	
Level not applicable Chemical code letter	Risks resist	tance resistance	res	istance resistance		
EN 374-1 chemical code letter				TDM-Test		
	Level not app	plicable				
	EN 374-1		mical			
Chemical hisks	Circumcar rasks	Sodium hydroxide 40%				ISO 374-1/Type B
		Hydrogen Peroxide 30%			T T	
Δ /	LIV 37 4 4	Formaldehyde 37%			1	
Degradation	Degradation					
KPT						КРТ
	EN 374-5 The glove	e is tight against microorganis	sms (viral, bacteria and fung	gi). Test according to ISO 16604	- method B.	
	microorganism			-		(ゑ)
microorganism	41-64					VIRUS
	tigntness		cording to EN ISO 21/20			
microorganism tightness	- Table 1	e meets the requirements acc	JUI UITIE LU EIN IOU Z142U			
microorganism tightness  EN ISO 21420 The glove meets the requirements according to EN ISO 21420	- Table 1	e meets the requirements acc	Loruing to EN 130 21420			
microorganism tightness  EN ISO 21420 protective gloves  The glove meets the requirements according to EN ISO 21420	EN ISO 21420 The glove protective gloves					
microorganism tightness  EN ISO 21420 protective gloves  The glove meets the requirements according to EN ISO 21420  EN 455  The glove meets the requirements according to EN 455-1, EN 455-2, EN 455-3, EN 455-4. Force at break ≥ 3.6 N.	EN ISO 21420 The glove protective gloves  EN 455 The glove			-2, EN 455-3, EN 455-4. Force at	break ≥ 3.6 N.	EN
microorganism tightness  EN ISO 21420 The glove meets the requirements according to EN ISO 21420 protective gloves  EN 455 The glove meets the requirements according to EN 455-1, EN 455-2, EN 455-3, EN 455-4. Force at break ≥ 3.6 N.  EN 455 The glove meets the requirements according to EN 455-1, EN 455-2, EN 455-3, EN 455-4. Force at break ≥ 3.6 N.	EN ISO 21420 The glove protective gloves  EN 455 The glove medical gloves for			-2, EN 455-3, EN 455-4. Force at	: break ≥ 3.6 N.	
microorganism tightness  EN ISO 21420 The glove meets the requirements according to EN ISO 21420 protective gloves  EN 455 The glove meets the requirements according to EN 455-1, EN 455-2, EN 455-3, EN 455-4. Force at break ≥ 3.6 N. EN 455 single use	EN ISO 21420 The glove protective gloves  EN 455 The glove medical gloves for single use	e meets the requirements acc	cording to EN 455-1, EN 455			455
microorganism tightness  EN ISO 21420 protective gloves  The glove meets the requirements according to EN ISO 21420  EN 455 medical gloves for single use  EN 455-1 The glove has an AQL < 1.5 in regards to the water retention test (sampling inspection in acc. to ISO 2859-1,	EN ISO 21420 The glove protective gloves  EN 455 The glove medical gloves for single use EN 455-1 The glove	e meets the requirements acc e has an AQL < 1.5 in regards	cording to EN 455-1, EN 455			<b>455</b>
microorganism tightness  EN ISO 21420 protective gloves  The glove meets the requirements according to EN ISO 21420  EN 455 medical gloves for single use  EN 455-1  The glove has an AQL < 1.5 in regards to the water retention test (sampling inspection in acc. to ISO 2859-1,	EN ISO 21420 The glove protective gloves  EN 455 The glove medical gloves for single use EN 455-1 The glove	e meets the requirements acc e has an AQL < 1.5 in regards	cording to EN 455-1, EN 455			<b>455</b>
microorganism tightness  EN ISO 21420 protective gloves  The glove meets the requirements according to EN ISO 21420  EN 455 medical gloves for single use EN 455-1 Fireedom from holes  The glove has an AQL < 1.5 in regards to the water retention test (sampling inspection in acc. to ISO 2859-1, general Inspection Level I)  AQL 1.5	EN ISO 21420 protective gloves  EN 455 medical gloves for single use EN 455-1 freedom from holes  The glove general lie	e meets the requirements acc e has an AQL < 1.5 in regards nspection Level I)	cording to EN 455-1, EN 455			<b>455</b>
microorganism tightness  EN ISO 21420 protective gloves  The glove meets the requirements according to EN ISO 21420  EN 455 medical gloves for single use EN 455-1 Fireedom from holes  The glove has an AQL < 1.5 in regards to the water retention test (sampling inspection in acc. to ISO 2859-1, general Inspection Level I)  EN 16350  not applicable	EN ISO 21420 protective gloves  EN 455 medical gloves for single use EN 455-1 freedom from holes  EN 16350 The glove gloves The glove general in	e meets the requirements acc e has an AQL < 1.5 in regards nspection Level I)	cording to EN 455-1, EN 455			<b>455</b>
microorganism tightness  EN ISO 21420 protective gloves  The glove meets the requirements according to EN ISO 21420  EN 455 medical gloves for single use EN 455-1 Fireedom from holes  The glove has an AQL < 1.5 in regards to the water retention test (sampling inspection in acc. to ISO 2859-1, general Inspection Level I)  EN 16350 electrostatic	EN ISO 21420 protective gloves  EN 455 medical gloves for single use EN 455-1 freedom from holes  The glove general lie	e meets the requirements acc e has an AQL < 1.5 in regards nspection Level I)	cording to EN 455-1, EN 455			<b>455</b>

QMFORM\_60.003 1/3 issue date: 16.04.2025









**4M**Pri

Article-No.: 01185

Description: **STYLE STRAWBERRY** 

Nitrile examination glove

light pink, non sterile, powder free

generell information	
material	carton
pieces per subpacking	100
GTIN subpacking size XS	4044941008868
GTIN subpacking size S	4044941008875
GTIN subpacking size M	4044941008882
GTIN subpacking size L	4044941008899
GTIN subpacking size XL	-
GTIN subpacking size XXL	-
GTIN subpacking size XXXL	-
PZN subpacking size XS	14819927
PZN subpacking size S	14819910
PZN subpacking size M	14819904
PZN subpacking size L	14819873
PZN subpacking size XL	-
PZN subpacking size XXL	-
PZN subpacking size XXXL	-
measures & size	
length	215 mm
width	120 mm
heigth	60 mm
weights	
size	gross weight*
XS	380 g
S	410 g
M	430 g
L	460 g
XL	-
XXL	-
XXXL	-

LOGISTIC DATA PALETTE	
general information	
kind of palett	euro-palette
measures & size	
cartons per layer	10
layers per palette	8
heigth of the palette	199 cm
weights	
size	gross weight*
XS	369 kg
S	393 kg
М	409 kg
L	433 kg
XL	-
XXL	-
XXXL	-

EN



generell information		
material	carton	
subpackings per outer packing	10	
GTIN outer packing size XS	4044941008912	
GTIN outer packing size S	4044941008929	
GTIN outer packing size M	4044941008936	
GTIN outer packing size L	4044941008943	
GTIN outer packing size XL	-	
GTIN outer packing size XXL	-	
GTIN outer packing size XXXL	-	
PZN outer packing size XS	-	
PZN outer packing size S	-	
PZN outer packing size M	-	
PZN outer packing size L	-	
PZN outer packing size XL	-	
PZN outer packing size XXL	-	
PZN outer packing size XXXL	-	
measures & size		
length	315 mm	
width	255 mm	
heigth	230 mm	
weights		
size	gross weight*	
XS	4.300 g	
S	4.600 g	
M	4.800 g	
L	5.100 g	
XL	-	
XXL	-	
XXXL	-	



## AMPri Handelsgesellschaft mbH

Benzstraße 16

21423 Winsen (Luhe) - Germany Telefon: +49 (0)4171 / 8480-0

> Homepage: www.ampri.de e-mail: info@ampri.de

## **Technical Data Sheet**

Article-No.: 01185

Description: STYLE STRAWBERRY

Nitrile examination glove

light pink, non sterile, powder free



### WARNINGS AND SAFETY INFORMATION

# storage / expiry date

Store gloves in original packaging in a cool and dry place without additional weight, protect from direct sunlight. Do not store near ozone sources (laser printers, copiers). The actual expiry time in use cannot be specified in general terms, as it depends on the general conditions of use. An individual risk assessment must be carried out in each case. The expiry date - valid for proper storage - is stated on the packaging.

#### use and control

Always use protective gloves only for the intended use and in the correct size. A check/risk assessment must be carried out to ensure that the gloves are suitable for the intended use, as the conditions at the workplace may deviate from those of the type test depending on temperature, abrasion and degradation. Breakthrough times and permeation levels are based on laboratory measurements and are determined using samples taken from the palm of the hand. The actual duration of protection of a glove with a specific substance can vary significantly due to the conditions of use (temperature, abrasion, stretching). In the case of aggressive chemicals, degradation (change in mechanical properties) can be an important factor to consider when selecting chemical-resistant gloves. This information does not reflect the actual duration of protection in the workplace and the distinction between mixtures and pure chemicals. The chemical resistance was determined under laboratory conditions only on the basis of samples from the palm and refers only to the chemicals tested. The situation may be different if the chemical is used in a mixture. The penetration resistance was evaluated under laboratory conditions and refers only to the tested specimen. The degradation results according to EN ISO 374-4 show the change in puncture resistance of the gloves after exposure to the tested chemical.

Before use, the gloves must be checked for holes or damage.

#### disposal

Used gloves must be disposed of in accordance with the disposal regulations of the local waste disposal company. Unused gloves can be disposed of with household waste.

## disinfection

Disinfection is not intended for these gloves and is the responsibility of the user.

#### warnings/ allergy information

Protective gloves are intended for single use only.

This product contains dithiocarbamates, which may cause allergic reactions

# donning and doffing instructions











\*slight deviations possible due to standard tolerances

rev-no.: 2025-01 date 16.06.2025

changes and errors excepted

QMFORM\_60.003 issue date: 16.04.2025