

21423 Winsen (Luhe) - Germany

Telefon: +49 (0)4171 / 8480-0 Homepage: www.ampri.de e-mail: info@ampri.de


# **Technical Data Sheet**

**4M**Pri

Article-No.: 01034

**MED-COMFORT Polymer Plus** Description:

Latex examination glove white, non sterile, powder free





PRODUCT DESCRIP	TION						
material		Nitrile	□ Vinyl	☐ Vinyl-Nitrile-	☐ Polyethy-lene	☐ TPE	☐ cotton
				mixture	(PE)		
colour	✓ white	□ blue		☐ mint	□ purple	□ mix	
characteristics	☐ prepowdered	powderfree	sterile	non sterile	☑ ambidex-	☐ fits hand-	☐ biodgra-
	_				trous	specific	dable
surface	☐ full textured	☐ finger	□ not textured	embossed	□ chlorinated ins	side	
		textured					
			I.				
SIZES	) (T. c)	0 (0 =)	22 (7.0)	. (0.0)	VII (0.40)	100 (40 44)	) non (44 40)
	XS (5-6)	S (6-7)	M (7-8)	L (8-9)	XL (9-10)	XXL (10-11)	XXXL (11-12)
width length	≤ 80 mm ≥ 240 mm	80 ± 10 mm ≥ 240 mm	95 ± 10 mm ≥ 240 mm	110 ± 10 mm ≥ 240 mm	115 ± 10 mm ≥ 240 mm	-	-
length	2 240 111111	2 240 111111	2 240 111111	2 240 111111	2 240 111111	-	-
REGULATORY AFFA	IRS						
PPE-Regulation	☐ Category I	☐ Category II	Category III	☐ no PPE-article			
(EU) 2016/425							( (
MD-Regulation	☑ Class I	Class II a	Class III	□ sterile	☐ measuring	no medical	$\epsilon$
(EU) 2017/745					function	device	
Food Contact	☑ acidic foods	☑ aqueous	☐ fatty foods	☑ alcoholic	☑ dry foods	☐ not approved	77
(EG) 1935/2004		foods		foods		for food-	501
						contact	
STANDARDISATION	J						
EN 388 Mechanical	abrasion	blade cut	tear resistance	puncture	blade cut	impact test	
Risks	resistance	resistance		resistance	resistance		
		Coupe-Test			TDM-Test		
Level	not applicable						
							I
FN 374-1	chemical		code letter	level	permeation time	degradation	
	<b>chemical</b> Sodium hydroxide	40%	code letter K	level 6	> 480 min	degradation -42,9 %	ISO 374-1/Type B
EN 374-1 Chemical Risks					•		ISO 374-1/Type B
	Sodium hydroxide	2 30%	K	6	> 480 min	-42,9 %	ISO 374-1/Type B
Chemical Risks	Sodium hydroxide Hydrogen Peroxide	2 30%	K P	6 5	> 480 min > 240 min	-42,9 % 22,8 %	ISO 374-1/Type B
Chemical Risks EN 374-4	Sodium hydroxide Hydrogen Peroxide	2 30%	K P	6 5	> 480 min > 240 min	-42,9 % 22,8 %	ISO 374-1/Type B
Chemical Risks EN 374-4 Degradation	Sodium hydroxide Hydrogen Peroxide Formaldehyde 37%	2 30%	K P T	6 5 2	> 480 min > 240 min > 30 min	-42,9 % 22,8 % 5,0 %	ISO 374-1/Type B
Chemical Risks EN 374-4 Degradation EN 374-5	Sodium hydroxide Hydrogen Peroxide Formaldehyde 37%	2 30%	K P T	6 5 2	> 480 min > 240 min	-42,9 % 22,8 % 5,0 %	KPT EN ISO 374-5:2016
Chemical Risks EN 374-4 Degradation EN 374-5 microorganism	Sodium hydroxide Hydrogen Peroxide Formaldehyde 37%	2 30%	K P T	6 5 2	> 480 min > 240 min > 30 min	-42,9 % 22,8 % 5,0 %	KPT  8N 150 324-5-2016
Chemical Risks EN 374-4 Degradation EN 374-5	Sodium hydroxide Hydrogen Peroxide Formaldehyde 37%	2 30%	K P T	6 5 2	> 480 min > 240 min > 30 min	-42,9 % 22,8 % 5,0 %	KPT EN ISO 374-5:2016
Chemical Risks EN 374-4 Degradation EN 374-5 microorganism	Sodium hydroxide Hydrogen Peroxide Formaldehyde 37% The glove is tight a	gainst microorganis	K P T	6 5 2 and fungi). Test acco	> 480 min > 240 min > 30 min	-42,9 % 22,8 % 5,0 %	KPT 8N 150 375-52016
Chemical Risks EN 374-4 Degradation EN 374-5 microorganism tightness	Sodium hydroxide Hydrogen Peroxide Formaldehyde 37% The glove is tight a	gainst microorganis	K P T T sms (viral, bacteria a	6 5 2 and fungi). Test acco	> 480 min > 240 min > 30 min	-42,9 % 22,8 % 5,0 %	KPT 8N 150 374-52016
Chemical Risks EN 374-4 Degradation EN 374-5 microorganism tightness EN ISO 21420 protective gloves	Sodium hydroxide Hydrogen Peroxide Formaldehyde 37%  The glove is tight a  The glove meets th	gainst microorganis	K P T  sms (viral, bacteria a	6 5 2 and fungi). Test acco	> 480 min > 240 min > 30 min > 30 min ording to ISO 16604	-42,9 % 22,8 % 5,0 %	KPT 8N 150 374-52016
Chemical Risks EN 374-4 Degradation EN 374-5 microorganism tightness EN ISO 21420 protective gloves EN 455	Sodium hydroxide Hydrogen Peroxide Formaldehyde 37%  The glove is tight a  The glove meets th	gainst microorganis	K P T  sms (viral, bacteria a	6 5 2 and fungi). Test acco	> 480 min > 240 min > 30 min > 30 min ording to ISO 16604	-42,9 % 22,8 % 5,0 %	KPT 8N 150 375-52016
Chemical Risks EN 374-4 Degradation EN 374-5 microorganism tightness EN ISO 21420 protective gloves EN 455 medical gloves for	Sodium hydroxide Hydrogen Peroxide Formaldehyde 37%  The glove is tight a  The glove meets th	gainst microorganis	K P T  sms (viral, bacteria a	6 5 2 and fungi). Test acco	> 480 min > 240 min > 30 min > 30 min ording to ISO 16604	-42,9 % 22,8 % 5,0 %	KPT EN 150 324 5-2016 VIRUS
Chemical Risks EN 374-4 Degradation EN 374-5 microorganism tightness EN ISO 21420 protective gloves EN 455 medical gloves for single use	Sodium hydroxide Hydrogen Peroxide Formaldehyde 37%  The glove is tight a  The glove meets th	gainst microorganis requirements acc	K P T  sms (viral, bacteria according to EN ISO 21	6 5 2 and fungi). Test acco	> 480 min > 240 min > 30 min > 30 min ording to ISO 16604	-42,9 % 22,8 % 5,0 %	KPT  EN ISO 2016  VIRUS  EN 455
Chemical Risks EN 374-4 Degradation EN 374-5 microorganism tightness EN ISO 21420 protective gloves EN 455 medical gloves for single use EN 455-1	Sodium hydroxide Hydrogen Peroxide Formaldehyde 37%  The glove is tight a  The glove meets th  The glove meets th	gainst microorganis requirements accorder requirements accorder requirements accorder accorde	K P T  sms (viral, bacteria according to EN ISO 21	6 5 2 and fungi). Test acco	> 480 min > 240 min > 30 min > 30 min ording to ISO 16604	-42,9 % 22,8 % 5,0 %	KPT  SN 150 2016-2016  VIRUS  EN 455
Chemical Risks EN 374-4 Degradation EN 374-5 microorganism tightness EN ISO 21420 protective gloves EN 455 medical gloves for single use	Sodium hydroxide Hydrogen Peroxide Formaldehyde 37%  The glove is tight a  The glove meets th	gainst microorganis requirements accorder requirements accorder requirements accorder accorde	K P T  sms (viral, bacteria according to EN ISO 21	6 5 2 and fungi). Test acco	> 480 min > 240 min > 30 min > 30 min ording to ISO 16604	-42,9 % 22,8 % 5,0 %	KPT  EN 150 JN-3-2016  VIRUS  EN 455
Chemical Risks EN 374-4 Degradation EN 374-5 microorganism tightness EN ISO 21420 protective gloves EN 455 medical gloves for single use EN 455-1 freedom from holes	Sodium hydroxide Hydrogen Peroxide Formaldehyde 37%  The glove is tight a  The glove meets th  The glove meets th  The glove has an A general Inspection	gainst microorganis requirements accorder requirements accorder requirements accorder accorde	K P T  sms (viral, bacteria according to EN ISO 21	6 5 2 and fungi). Test acco	> 480 min > 240 min > 30 min > 30 min ording to ISO 16604	-42,9 % 22,8 % 5,0 %	KPT  KPT  KPT  KPT  KPT  KPT  KPT  KPT
Chemical Risks EN 374-4 Degradation EN 374-5 microorganism tightness EN ISO 21420 protective gloves EN 455 medical gloves for single use EN 455-1 freedom from holes	Sodium hydroxide Hydrogen Peroxide Formaldehyde 37%  The glove is tight a  The glove meets th  The glove meets th	gainst microorganis requirements accorder requirements accorder requirements accorder accorde	K P T  sms (viral, bacteria according to EN ISO 21	6 5 2 and fungi). Test acco	> 480 min > 240 min > 30 min > 30 min ording to ISO 16604	-42,9 % 22,8 % 5,0 %	KPT  N 150 274-52916  VIRUS  EN  455
Chemical Risks EN 374-4 Degradation EN 374-5 microorganism tightness EN ISO 21420 protective gloves EN 455 medical gloves for single use EN 455-1 freedom from holes	Sodium hydroxide Hydrogen Peroxide Formaldehyde 37%  The glove is tight a  The glove meets th  The glove meets th  The glove has an A general Inspection	gainst microorganis requirements accorder requirements accorder requirements accorder accorde	K P T  sms (viral, bacteria according to EN ISO 21	6 5 2 and fungi). Test acco	> 480 min > 240 min > 30 min > 30 min ording to ISO 16604	-42,9 % 22,8 % 5,0 %	KPT  N 150 278-52016  VIRUS  EN  455
Chemical Risks EN 374-4 Degradation  EN 374-5 microorganism tightness EN ISO 21420 protective gloves EN 455 medical gloves for single use EN 455-1 freedom from holes	Sodium hydroxide Hydrogen Peroxide Formaldehyde 37%  The glove is tight a  The glove meets th  The glove meets th  The glove has an A general Inspection	gainst microorganis requirements accorder requirements accorder requirements accorder accorde	K P T  sms (viral, bacteria according to EN ISO 21	6 5 2 and fungi). Test acco	> 480 min > 240 min > 30 min > 30 min ording to ISO 16604	-42,9 % 22,8 % 5,0 %	KPT  N 150 278-52016  VIRUS  EN  455



21423 Winsen (Luhe) - Germany Telefon: +49 (0)4171 / 8480-0 Homepage: www.ampri.de

e-mail: info@ampri.de

## **Technical Data Sheet**

Article-No.: 01034

Description: MED-COMFORT Polymer Plus

Latex examination glove white, non sterile, powder free

LOGISTIC DATA S		
generell informatio	1	
material	carton	
pieces per subpackir	100	
GTIN subpacking size	4044941000817	
GTIN subpacking size	4044941000824	
GTIN subpacking size	4044941000831	
GTIN subpacking size	4044941000848	
GTIN subpacking size	4044941000855	
GTIN subpacking size	-	
GTIN subpacking size	-	
PZN subpacking size	15591797	
PZN subpacking size	15591828	
PZN subpacking size	15591811	
PZN subpacking size L		15591805
PZN subpacking size XL		15591834
PZN subpacking size	=	
PZN subpacking size	-	
measures & size		
length	240 mm	
width	120 mm	
heigth	70 mm	
weights		
size	net weight	gross weight
XS	550 g	610 g
S	600 g	660 g
М	650 g	710 g
L	700 g	760 g
XL	750 g	810 g
XXL	-	-
XXXL	_	_

LOGISTIC DATA PALETTE				
general information	1			
kind of palett		euro-palette		
measures & size				
cartons per layer		9		
layers per palette		7		
heigth of the palette		189 cm		
weights				
size	net weight	gross weight		
XS	416 g	441 g		
S 447 g		472 g		
М	479 g	504 g		
L	510 g	535 g		
XL	542 g	567 g		
XXL	-	-		
XXXI	-	_		



LOGISTIC DATA	OUTER PACKING			
generell information				
material		carton		
subpackings per ou	ter packing	10		
GTIN outer packing	size XS	4044941002712		
GTIN outer packing	size S	4044941002729		
GTIN outer packing	size M	4044941002736		
GTIN outer packing	4044941002743			
GTIN outer packing	4044941002750			
GTIN outer packing	size XXL	-		
GTIN outer packing	size XXXL	-		
PZN outer packing	size XS	-		
PZN outer packing	size S	-		
PZN outer packing	size M	-		
PZN outer packing	-			
PZN outer packing	-			
PZN outer packing	size XXL	-		
PZN outer packing :	size XXXL	-		
measures & size				
length		260 mm		
width		260 mm		
heigth		248 mm		
weights				
size	net weight	gross weight		
XS	6.100 g	6.600 g		
S	6.600 g	7.100 g		
M	7.100 g	7.600 g		
L	7.600 g	8.100 g		
XL	8.100 g	8.600 g		
XXL	-	-		
XXXL	-			





## AMPri Handelsgesellschaft mbH

Benzstraße 16

21423 Winsen (Luhe) - Germany Telefon: +49 (0)4171 / 8480-0

Homepage: www.ampri.de e-mail: info@ampri.de

## **Technical Data Sheet**

**4M**Pri

Article-No.: 01034

Description: MED-COMFORT Polymer Plus

Latex examination glove white, non sterile, powder free



## WARNINGS AND SAFETY INFORMATION

# storage / expiry date

Store gloves in original packaging in a cool and dry place without additional weight, protect from direct sunlight. Do not store near ozone sources (laser printers, copiers). The actual expiry time in use cannot be specified in general terms, as it depends on the general conditions of use. An individual risk assessment must be carried out in each case. The expiry date - valid for proper storage - is stated on the packaging.

#### use and control

Always use protective gloves only for the intended use and in the correct size. A check/risk assessment must be carried out to ensure that the gloves are suitable for the intended use, as the conditions at the workplace may deviate from those of the type test depending on temperature, abrasion and degradation. Breakthrough times and permeation levels are based on laboratory measurements and are determined using samples taken from the palm of the hand. The actual duration of protection of a glove with a specific substance can vary significantly due to the conditions of use (temperature, abrasion, stretching). In the case of aggressive chemicals, degradation (change in mechanical properties) can be an important factor to consider when selecting chemical-resistant gloves. This information does not reflect the actual duration of protection in the workplace and the distinction between mixtures and pure chemicals. The chemical resistance was determined under laboratory conditions only on the basis of samples from the palm and refers only to the chemicals tested. The situation may be different if the chemical is used in a mixture. The penetration resistance was evaluated under laboratory conditions and refers only to the tested specimen. The degradation results according to EN ISO 374-4 show the change in puncture resistance of the gloves after exposure to the tested chemical.

Before use, the gloves must be checked for holes or damage.

## disposal

Used gloves must be disposed of after contact with chemicals in accordance with the disposal regulations for the chemical and the regulations of the local waste disposal company. Unused gloves can be disposed of with household waste.

### disinfection

Disinfection is not intended for these gloves and is the responsibility of the user.

#### warnings/ allergy information

Protective gloves are intended for single use only.

This product contains dithiocarbamates and natural latex, which can trigger allergic reactions, including anaphylactic reactions

# donning and doffing instructions











rev-no.: 10

date 09.09.2024 changes and errors excepted