

Benzstraße 16 21423 Winsen (Luhe) - Germany

Telefon: +49 (0)4171 / 8480-0

Homepage: www.ampri.de e-mail: info@ampri.de

Technical Data Sheet


Article-No.: 118-069

PRODUCT DESCRIPTION

EN

Description: **Nature Glove**

Nitrile examination glove green, non sterile, powder free





PRODUCT DESCR	IPTION						
material	Latex	✓ Nitrile	□ Vinyl	☐ Vinyl-Nitrile- mixture	Polyethy-lene (PE)	□ ТРЕ	cotton
colour	☐ white	□ blue	□ black	☐ mint	☐ purple	☐ mix	☑ green
characteristics	prepowdered		sterile	non sterile	✓ ambidex-	fits hand-	☑ biodegra-
					trous	specific	dable
surface	textured	not textured	embossed			- 1	
SIZES							
	XS (5-6)	S (6-7)	M (7-8)	L (8-9)	XL (9-10)	XXL (10-11)	XXXL (11-12)
width	≤ 80 mm	80 ± 10 mm	95 ± 10 mm	110 ± 10 mm	115 ± 10 mm	1	-
length	≥ 240 mm	≥ 240 mm	≥ 240 mm	≥ 240 mm	≥ 240 mm	-	-
REGULATORY AF	FAIRS						
PPE-Regulation	☐ Category I	☐ Category II	∇ Category III	☐ no PPE-article			
(EU) 2016/425							((
MD-Regulation	Class I	Class II a	Class III	sterile	measuring	no medical	CE
(EU) 2017/745					function	device	
Food Contact	acidic foods	aqueous	fatty foods	alcoholic	dry foods	not approved	ניור ו
(EG) 1935/2004		foods		foods		for food-	521
						contact	
STANDARDISATIO	ON						
EN 388 Mechanical	abrasion	blade cut	tear resistance	puncture	blade cut	impact test	
Risks	resistance	resistance		resistance	resistance		
		Coupe-Test			TDM-Test		
Level	not applicable						
EN 374-1		che	mical		code l	letter	
Chemical Risks	Sodium hydroxide	: 40%			K	(ISO 374-1/Type 0
	Formaldehyde 379	%			T	-	
EN 374-4							()()
Degradation							
							~
EN 374-5							
LIT 37 4 3	The glove is tight a	against microorgani	sms (viral hacteria	and fungi) Test acco	ording to ISO 16604	- method B	EN ISO 374-5:2016
microorganism	The glove is tight a	against microorgani	sms (viral, bacteria	and fungi). Test acco	ording to ISO 16604	- method B.	EN ISO 374-5:2016
_	The glove is tight a	against microorgani	sms (viral, bacteria	and fungi). Test acco	ording to ISO 16604	- method B.	EN ISO 374-5:2016
tightness					ording to ISO 16604	- method B.	
tightness EN ISO 21420		against microorganis			ording to ISO 16604	- method B.	
tightness EN ISO 21420					ording to ISO 16604	- method B.	
tightness EN ISO 21420 protective gloves	The glove meets to	he requirements acc	cording to EN ISO 2				VIRUS
tightness EN ISO 21420 protective gloves EN 455	The glove meets to	he requirements acc	cording to EN ISO 2	1420			EN EN
tightness EN ISO 21420 protective gloves EN 455 medical gloves for	The glove meets to	he requirements acc	cording to EN ISO 2	1420			VIRUS
tightness EN ISO 21420 protective gloves EN 455 medical gloves for single use	The glove meets to	he requirements acc	cording to EN ISO 2	1420	3, EN 455-4. Force a	t break ≥ 6 N.	EN 455
tightness EN ISO 21420 protective gloves EN 455 medical gloves for single use EN 455-1	The glove meets to The glove meets to The glove has an A	he requirements accomments accomm	cording to EN ISO 2	1420 I, EN 455-2, EN 455-	3, EN 455-4. Force a	t break ≥ 6 N.	EN 455
tightness EN ISO 21420 protective gloves EN 455 medical gloves for single use EN 455-1	The glove meets to The glove meets to The glove has an A	he requirements accomments accomm	cording to EN ISO 2	1420 I, EN 455-2, EN 455-	3, EN 455-4. Force a	t break ≥ 6 N.	EN 455
microorganism tightness EN ISO 21420 protective gloves EN 455 medical gloves for single use EN 455-1 freedom from holes ASTM D5526	The glove meets to The glove meets to The glove has an A general Inspection	he requirements accomments accomm	cording to EN ISO 2 cording to EN 455-1 to the water retent	1420 I, EN 455-2, EN 455-	3, EN 455-4. Force a inspection in acc. to	t break ≥ 6 N. ISO 2859-1,	EN 455
tightness EN ISO 21420 protective gloves EN 455 medical gloves for single use EN 455-1 freedom from holes	The glove meets to The glove meets to The glove has an A general Inspection The glove meets to	he requirements accomments accomm	cording to EN ISO 2 cording to EN 455-1 to the water retent	1420 I, EN 455-2, EN 455- tion test (sampling	3, EN 455-4. Force a inspection in acc. to	t break ≥ 6 N. ISO 2859-1,	EN 455

QMFORM 60.003 1/3





Benzstraße 16 21423 Winsen (Luhe) - Germany

Telefon: +49 (0)4171 / 8480-0 Homepage: www.ampri.de

e-mail: info@ampri.de

Technical Data Sheet

Article-No.: **118-069**

Description: Nature Glove

Nitrile examination glove green, non sterile, powder free

LOGISTIC DATA SUBPACKING	
generell information	
material	carton
pieces per subpacking	100
GTIN subpacking size XS	4044941712857
GTIN subpacking size S	4044941712307
GTIN subpacking size M	4044941712321
GTIN subpacking size L	4044941712345
GTIN subpacking size XL	4044941712396
GTIN subpacking size XXL	-
GTIN subpacking size XXXL	-
PZN subpacking size XS	18034169
PZN subpacking size S	18034152
PZN subpacking size M	18034146
PZN subpacking size L	18034123
PZN subpacking size XL	18034117
PZN subpacking size XXL	-
PZN subpacking size XXXL	-
measures & size	
length	215 mm
width	120 mm
heigth	60 mm
weights	
size	gross weight*
XS	380 g
S	410 g
M	440 g
L	460 g
XL	500 g
XXL	-
XXXL	-

LOGISTIC DATA PALETTE	
general information	
kind of palett	euro-palette
measures & size	
cartons per layer	10
layers per palette	8
heigth of the palette	199 cm
weights	
size	gross weight*
XS	369 kg
S	393 kg
М	417 kg
L	433 kg
XL	465 kg
XXL	-
XXXL	-

ΕN



LOGISTIC DATA OUTER PACKING			
generell information			
material	carton		
subpackings per outer packing	10		
GTIN outer packing size XS	4044941712864		
GTIN outer packing size S	4044941712314		
GTIN outer packing size M	4044941712338		
GTIN outer packing size L	4044941712352		
GTIN outer packing size XL	4044941712376		
GTIN outer packing size XXL	-		
GTIN outer packing size XXXL	-		
PZN outer packing size XS	-		
PZN outer packing size S	-		
PZN outer packing size M	-		
PZN outer packing size L	-		
PZN outer packing size XL	-		
PZN outer packing size XXL	-		
PZN outer packing size XXXL	-		
measures & size			
length	315 mm		
width	255 mm		
heigth	230 mm		
weights			
size	gross weight*		
XS	4.300 g		
S	4.600 g		
M	4.900 g		
L	5.100 g		
XL	5.500 g		
XXL	-		
XXXL	-		



AMPri Handelsgesellschaft mbH

Benzstraße 16

21423 Winsen (Luhe) - Germany

Telefon: +49 (0)4171 / 8480-0 Homepage: www.ampri.de e-mail: info@ampri.de

Technical Data Sheet

118-069 Article-No.:

Description: **Nature Glove**

> Nitrile examination glove green, non sterile, powder free



WARNINGS AND SAFETY INFORMATION

storage / expiry date

Store gloves in original packaging in a cool and dry place without additional weight, protect from direct sunlight. Do not store near ozone sources (laser printers, copiers). The actual expiry time in use cannot be specified in general terms, as it depends on the general conditions of use. An individual risk assessment must be carried out in each case. The expiry date - valid for proper storage - is stated on

use and control

Always use protective gloves only for the intended use and in the correct size. A check/risk assessment must be carried out to ensure that the gloves are suitable for the intended use, as the conditions at the workplace may deviate from those of the type test depending on temperature, abrasion and degradation. Breakthrough times and permeation levels are based on laboratory measurements and are determined using samples taken from the palm of the hand. The actual duration of protection of a glove with a specific substance can vary significantly due to the conditions of use (temperature, abrasion, stretching). In the case of aggressive chemicals, degradation (change in mechanical properties) can be an important factor to consider when selecting chemical-resistant gloves. This information does not reflect the actual duration of protection in the workplace and the distinction between mixtures and pure chemicals. The chemical resistance was determined under laboratory conditions only on the basis of samples from the palm and refers only to the chemicals tested. The situation may be different if the chemical is used in a mixture. The penetration resistance was evaluated under laboratory conditions and refers only to the tested specimen. The degradation results according to EN ISO 374-4 show the change in puncture resistance of the gloves after exposure to the tested chemical. Before use, the gloves must be checked for holes or damage.

disposal

Used gloves must be disposed of after contact with chemicals in accordance with the disposal regulations for the chemical and the regulations of the local waste disposal company. Unused gloves can be disposed of with household waste.

disinfection

Disinfection is not intended for these gloves and is the responsibility of the user.

warnings/ allergy information

Protective gloves are intended for single use only.

This product contains dithiocarbamates, which may cause allergic reactions

donning and doffing instructions











*slight deviations possible due to standard tolerances

2025-02 07.11.2025 date

changes and errors excepted