

Telefon: +49 (0)4171 / 8480-0

Homepage: www.ampri.de e-mail: info@ampri.de

Technical Data Sheet


4MPri

Article-No.: 01183

0,06 mm

Description:

0,08 mm



0,10 mm



PRODUCT DESCRIP	TION						
material	Latex	✓ Nitrile	☐ Vinyl	☐ Vinyl-Nitrile-	☐ Polyethy-lene	☐ TPE	□ cotton
				mixture	(PE)		
colour	□ white	blue	black	□ mint	□ purple	□ mix	☑ dark-blue
characteristics	☐ prepowdered	powderfree	☐ sterile	non sterile	☑ ambidex-	fits hand-	☐ biodegra-
					trous	specific	dable
surface		☐ not textured	□ embossed				
SIZES							
	XS (5-6)	S (6-7)	M (7-8)	L (8-9)	XL (9-10)	XXL (10-11)	XXXL (11-12)
width	≤ 80 mm	80 ± 10 mm	95 ± 10 mm	110 ± 10 mm	115 ± 10 mm	125 ± 10 mm	-
length	≥ 240 mm	≥ 240 mm	≥ 240 mm	≥ 240 mm	≥ 240 mm	≥ 240 mm	-
REGULATORY AFFA	AIRS						
PPE-Regulation	☐ Category I	☐ Category II	✓ Category III	no PPE-article			
(EU) 2016/425	Category	category ii	category iii	norre urticle			
MD-Regulation	☑ Class I	☐ Class II a	☐ Class III	□ sterile	☐ measuring	no medical	CE
(EU) 2017/745					function	device	, ,
Food Contact	☑ acidic foods	☑ aqueous	✓ fatty foods	☑ alcoholic	☑ dry foods	☐ not approved	
(EG) 1935/2004		foods		foods		for food-	5211
						contact	
STANDARDISATION	N						
EN 388 Mechanical	abrasion	blade cut	tear resistance	puncture	blade cut	impact test	
Risks				•			
NISKS	resistance	resistance		resistance	resistance		
NISKS	resistance	resistance Coupe-Test		resistance	resistance TDM-Test		
Level	not applicable			resistance			
	not applicable	Coupe-Test cher	mical	resistance	TDM-Test		
Level	not applicable Sodium hydroxide	Coupe-Test cher	nical	resistance	TDM-Test code	(ISO 374-1/Type B
Level EN 374-1	not applicable Sodium hydroxide Hydrogen Peroxide	Coupe-Test cher 40% 2 30%	nical	resistance	TDM-Test code	()	ISO 374-1/Type B
EN 374-1 Chemical Risks	not applicable Sodium hydroxide	Coupe-Test cher 40% 2 30%	mical	resistance	TDM-Test code	()	ISO 374-1/Type B
Level EN 374-1 Chemical Risks	not applicable Sodium hydroxide Hydrogen Peroxide	Coupe-Test cher 40% 2 30%	mical	resistance	TDM-Test code	()	ISO 374-1/Type B
EN 374-1 Chemical Risks	not applicable Sodium hydroxide Hydrogen Peroxide	Coupe-Test cher 40% 2 30%	nical	resistance	TDM-Test code	()	
EN 374-1 Chemical Risks	not applicable Sodium hydroxide Hydrogen Peroxidt Formaldehyde 379	Coupe-Test chei			Code		KPT 8N 150 274-52016
Level EN 374-1 Chemical Risks EN 374-4 Degradation	not applicable Sodium hydroxide Hydrogen Peroxidt Formaldehyde 379	Coupe-Test chei			TDM-Test code		KPT EN 150 274-5:2016
EN 374-1 Chemical Risks EN 374-4 Degradation	not applicable Sodium hydroxide Hydrogen Peroxidt Formaldehyde 379	Coupe-Test chei			Code		KPT
EN 374-1 Chemical Risks EN 374-4 Degradation EN 374-5 microorganism	not applicable Sodium hydroxide Hydrogen Peroxide Formaldehyde 379 The glove is tight a	cher 40% ≥ 30% 6 gainst microorganis	ims (viral, bacteria a	and fungi). Test acco	Code		KPT EN 150 274-5:2016
Level EN 374-1 Chemical Risks EN 374-4 Degradation EN 374-5 microorganism tightness	not applicable Sodium hydroxide Hydrogen Peroxide Formaldehyde 379 The glove is tight a	Coupe-Test chei	ims (viral, bacteria a	and fungi). Test acco	Code		KPT EN 150 274-5:2016
EN 374-1 Chemical Risks EN 374-4 Degradation EN 374-5 microorganism tightness EN ISO 21420 protective gloves	not applicable Sodium hydroxide Hydrogen Peroxide Formaldehyde 37% The glove is tight a	cher 40% 2 30% 6 gainst microorganis	ims (viral, bacteria a	and fungi). Test acco	rding to ISO 16604 -	method B.	KPT (N 150 274-52016
Level EN 374-1 Chemical Risks EN 374-4 Degradation EN 374-5 microorganism tightness EN ISO 21420 protective gloves EN 455	not applicable Sodium hydroxide Hydrogen Peroxide Formaldehyde 37% The glove is tight a	cher 40% 2 30% 6 gainst microorganis	ims (viral, bacteria a	and fungi). Test acco	Code	method B.	KPT EN 150 274-5:2016
Level EN 374-1 Chemical Risks EN 374-4 Degradation EN 374-5 microorganism tightness EN ISO 21420 protective gloves EN 455 medical gloves for	not applicable Sodium hydroxide Hydrogen Peroxide Formaldehyde 37% The glove is tight a	cher 40% 2 30% 6 gainst microorganis	ims (viral, bacteria a	and fungi). Test acco	rding to ISO 16604 -	method B.	KPT EN 150 274-5:2016 VIRUS
Level EN 374-1 Chemical Risks EN 374-4 Degradation EN 374-5 microorganism tightness EN ISO 21420 protective gloves EN 455 medical gloves for single use	not applicable Sodium hydroxide Hydrogen Peroxide Formaldehyde 379 The glove is tight a The glove meets th	cher 40% 2 30% 6 gainst microorganis ne requirements accome requirements accome	ording to EN ISO 21	and fungi). Test acco	rding to ISO 16604 -	method B.	KPT 18 150 324-52016 WIRUS EN 455
Level EN 374-1 Chemical Risks EN 374-4 Degradation EN 374-5 microorganism tightness EN ISO 21420 protective gloves EN 455 medical gloves for	not applicable Sodium hydroxide Hydrogen Peroxide Formaldehyde 379 The glove is tight a The glove meets th The glove meets th	cher 40% 2 30% 6 gainst microorganis ne requirements acco	ording to EN ISO 21	and fungi). Test acco	rding to ISO 16604 -	method B.	KPT EN ISO 324-52916 VIRUS EN 455
Level EN 374-1 Chemical Risks EN 374-4 Degradation EN 374-5 microorganism tightness EN ISO 21420 protective gloves EN 455 medical gloves for single use EN 455-1	not applicable Sodium hydroxide Hydrogen Peroxide Formaldehyde 379 The glove is tight a The glove meets th	cher 40% 2 30% 6 gainst microorganis ne requirements acco	ording to EN ISO 21	and fungi). Test acco	rding to ISO 16604 -	method B.	KPT EN 150 324-52016 WIRUS EN 455
Level EN 374-1 Chemical Risks EN 374-4 Degradation EN 374-5 microorganism tightness EN ISO 21420 protective gloves EN 455 medical gloves for single use EN 455-1 freedom from holes	not applicable Sodium hydroxide Hydrogen Peroxide Formaldehyde 379 The glove is tight a The glove meets th The glove meets th The glove has an A general Inspection	cher 40% 2 30% 6 gainst microorganis ne requirements acco	ording to EN ISO 21	and fungi). Test acco	rding to ISO 16604 -	method B.	KPT EN ISO 324-52916 VIRUS EN 455
Level EN 374-1 Chemical Risks EN 374-4 Degradation EN 374-5 microorganism tightness EN ISO 21420 protective gloves EN 455 medical gloves for single use EN 455-1	not applicable Sodium hydroxide Hydrogen Peroxide Formaldehyde 379 The glove is tight a The glove meets th The glove meets th	cher 40% 2 30% 6 gainst microorganis ne requirements acco	ording to EN ISO 21	and fungi). Test acco	rding to ISO 16604 -	method B.	KPT EN ISO 324-52916 VIRUS EN 455
Level EN 374-1 Chemical Risks EN 374-4 Degradation EN 374-5 microorganism tightness EN ISO 21420 protective gloves EN 455 medical gloves for single use EN 455-1 freedom from holes	not applicable Sodium hydroxide Hydrogen Peroxide Formaldehyde 379 The glove is tight a The glove meets th The glove meets th The glove has an A general Inspection	cher 40% 2 30% 6 gainst microorganis ne requirements acco	ording to EN ISO 21	and fungi). Test acco	rding to ISO 16604 -	method B.	KPT EN ISO 324-52916 VIRUS EN 455

QMFORM_60.003 issue date: 16.04.2025



21423 Winsen (Luhe) - Germany Telefon: +49 (0)4171 / 8480-0

Homepage: www.ampri.de e-mail: info@ampri.de


Technical Data Sheet

Article-No.: 01183

Description: STYLE BLUEBERRY

Nitrile examination glove

dark-blue, non sterile, powder free



LOGISTIC DATA PALETTE	
general information	
kind of palett	euro-palette
measures & size	
cartons per layer	10
layers per palette	8
heigth of the palette	199 cm
weights	
size	gross weight*
XS	369 kg
S	393 kg
М	409 kg
L	433 kg
XL	465 kg
XXL	473 kg
XXXL	-



LOGISTIC DATA OUTER PACKING	
generell information	
material	carton
subpackings per outer packing	10
GTIN outer packing size XS	4044941009216
GTIN outer packing size S	4044941009223
GTIN outer packing size M	4044941009230
GTIN outer packing size L	4044941009247
GTIN outer packing size XL	4044941009254
GTIN outer packing size XXL	4044941739953
GTIN outer packing size XXXL	-
PZN outer packing size XS	-
PZN outer packing size S	-
PZN outer packing size M	-
PZN outer packing size L	-
PZN outer packing size XL	-
PZN outer packing size XXL	-
PZN outer packing size XXXL	-
measures & size	
length	315 mm
width	255 mm
heigth	230 mm
weights	
size	gross weight*
XS	4.300 g
S	4.600 g
М	4.800 g
L	5.100 g
XL	5.500 g
XXL	5.600 g
XXXL	-



AMPri Handelsgesellschaft mbH

Benzstraße 16

21423 Winsen (Luhe) - Germany Telefon: +49 (0)4171 / 8480-0

Homepage: www.ampri.de e-mail: info@ampri.de

Technical Data Sheet

Article-No.: 01183

Description: STYLE BLUEBERRY

Nitrile examination glove

dark-blue, non sterile, powder free



WARNINGS AND SAFETY INFORMATION

storage /	exp	iry
date		

Store gloves in original packaging in a cool and dry place without additional weight, protect from direct sunlight. Do not store near ozone sources (laser printers, copiers). The actual expiry time in use cannot be specified in general terms, as it depends on the general conditions of use. An individual risk assessment must be carried out in each case. The expiry date - valid for proper storage - is stated on the packaging.

use and control

Always use protective gloves only for the intended use and in the correct size. A check/risk assessment must be carried out to ensure that the gloves are suitable for the intended use, as the conditions at the workplace may deviate from those of the type test depending on temperature, abrasion and degradation. Breakthrough times and permeation levels are based on laboratory measurements and are determined using samples taken from the palm of the hand. The actual duration of protection of a glove with a specific substance can vary significantly due to the conditions of use (temperature, abrasion, stretching). In the case of aggressive chemicals, degradation (change in mechanical properties) can be an important factor to consider when selecting chemical-resistant gloves. This information does not reflect the actual duration of protection in the workplace and the distinction between mixtures and pure chemicals. The chemical resistance was determined under laboratory conditions only on the basis of samples from the palm and refers only to the chemicals tested. The situation may be different if the chemical is used in a mixture. The penetration resistance was evaluated under laboratory conditions and refers only to the tested specimen. The degradation results according to EN ISO 374-4 show the change in puncture resistance of the gloves after exposure to the tested chemical.

Before use, the gloves must be checked for holes or damage.

disposal

Used gloves must be disposed of in accordance with the disposal regulations of the local waste disposal company. Unused gloves can be disposed of with household waste.

disinfection

Disinfection is not intended for these gloves and is the responsibility of the user.

warnings/ allergy information

Protective gloves are intended for single use only.

This product contains dithiocarbamates, which may cause allergic reactions

donning and doffing instructions











*slight deviations possible due to standard tolerances

rev-no.: 2025-03 date 07.11.2025

changes and errors excepted

QMFORM_60.003 issue date: 16.04.2025